

Before the
Federal Communications Commission
Washington, D.C. 20554

In the Matter of)
Investigation of the Spectrum Requirements for) ET Docket No. 06-135
Advanced Medical Technologies)
Amendment of Parts 2 and 95 of the) RM-11271
Commission's Rules to Establish the Medical)
Device Radiocommunication Service at)
401-402 and 405-406 MHz)
DexCom, Inc., Request for Waiver of the) ET Docket No. 05-213
Frequency Monitoring Requirements of the)
Medical Implant Communications Service Rules)
Biotronik, Inc., Request for Waiver of the) ET Docket No. 03-92
Frequency Monitoring Requirements of the)
Medical Implant Communications Service Rules)

ERRATUM

Released: May 6, 2009

By the Chief, Office of Engineering and Technology:

On March 20, 2009, the Commission released a Report and Order, FCC 09-23, in the above captioned proceeding. This Erratum corrects Appendix A of the Report and Order as indicated below:

- 1. Section 95.628 is corrected by deleting paragraph (g)(3)(ii).

* * * * *

- 2. Section 95.635(d)(6) is revised to read as follows:

(6) Emissions from a MedRadio transmitter operating in the 401-402 MHz or 405-406 MHz bands that are more than 100 kHz outside of either the 401-402 MHz or 405-406 MHz bands, and all emissions from such transmitter in the band 406.000-406.100 MHz shall be attenuated to a level no greater than the following field strength limits:

* * * * *

FEDERAL COMMUNICATIONS COMMISSION

Julius P. Knapp
Chief, Office of Engineering and Technology